

**CERTIFICATE OF COMPLIANCE**



**EL Class I**

*Date of issue: 3 March 2025*

*Valid until: 31 December 2026*

*EHEDG hereby declares that the product*

***Rosemount™ Sanitary Pressure Sensors, models 2051HT & 3051HT, with stainless isolator plugs type 2088-2801 or direct process connections type 2088- 2814 in sizes DN25, DN40, DN50, 1.5”, 2”***

*from*

*Rosemount Inc., 6021 Innovation Blvd, 55379 Shakopee, United States*

*has/have been evaluated for compliance and meets/meet the current criteria for Hygienic Equipment Design of closed process applications of the EHEDG*

***Certificate No. EHEDG-C2500007***

Signed \_\_\_\_\_ *Hein Timmerman* \_\_\_\_\_ *President EHEDG*

Signed \_\_\_\_\_ *Karlijn Faber* \_\_\_\_\_ *EHEDG Certification Officer*

*EHEDG  
Karspeldreef 8  
1101 CJ Amsterdam  
Netherlands*

©EHEDG



## APPENDIX 3 EHEDG Certification – Equipment Evaluation Form

Design Evaluation Date: 20.01.2025

EHEDG File Number: EHEDG-R2400026

Certification Type: EL CLASS I

Applicant: Rosemount, Inc.

Equipment: **Rosemount™ Sanitary Pressure Sensors, models 2051HT & 3051HT**, with stainless isolator plugs type 2088-2801 or direct process connections type 2088-2814 in sizes DN25, DN40, DN50, 1.5", 2".

Other essential identification:

**Evaluated by:**

Name: Mark Morgan, University of Tennessee

Date, Signature: 2/21/25 Mark T Morgan

**Approved by:**

Name: Andy Timperley, Chairman of SC Certification

Date, Signature: 27.02.2025, [Signature]

The use of the EHEDG Certification logo is justified based on the results of the design evaluation, inspection, and testing (as applicable) of the equipment for compliance with the current EHEDG Hygienic Design Criteria (HDC):

Criteria	Certification for <b>use in Closed Processes</b>
<input checked="" type="checkbox"/>	The equipment complies with all applicable HDC in the Guidelines.
<input type="checkbox"/>	Evidence for compliance required and provided by in-place cleanability test method according to EHEDG Doc. 2.
<input type="checkbox"/>	Evidence for compliance required and provided by in-place cleanability test method according to EHEDG Doc. 2, in-place sterilisability test method according to EHEDG Doc. 5, and bacteria tightness test according to EHEDG Doc. 7 for <b>EL ASEPTIC</b> Certification.
Criteria	Certification for <b>use in Open Processes</b>
<input type="checkbox"/>	The equipment complies with all applicable HDC in the Guidelines.
<input type="checkbox"/>	Evidence for compliance required and provided by OPC cleanability test method according to EHEDG Doc. 57.

## APPENDIX 3

No.	Description
1.	EHEDG Certificate of Conformance
2.	Contract to use the EHEDG Certification Logo for equipment
3.	Appendix 1: Equipment intended for cleaning in place with liquids without dismantling
4.	Appendix 2: Conditions for use of the EHEDG Certification logo
5.	Appendix 3: Equipment Evaluation form (this document)
6.	Evaluation Report for: <b>Rosemount™ Sanitary Pressure Sensors, models 2051HT &amp; 3051HT</b> , with stainless process connections type 2088-2814 (DN25, DN40, DN50, 1.5", 2") or stainless isolator plugs type 2088-2801 (1.5", 2"), report No. UTE-012025
7.	Drawings of the direct process connections and isolator plugs for 2051HT and 3051HT sanitary pressure sensors, DWG. No. 2088-2801 Rev AJ sheets 1,2,5,6,9,10,13,14 included (missing sheets are similar except different size of each type); 2088-2814 Rev AA Sheet 1 of 8 (included to show wetted surface similar to 2801 model connections), original stamped.
8.	Cleaning and Installation instructions, excerpts from Quick Start Guide 00825-01-4091, Rev DD October 2023 supplied by the manufacturer.
9.	Example of EHEDG Certified Logo Type EL Class I