

CERTIFICATE OF COMPLIANCE



EL Class I

Date of issue: 17 April 2025

Valid until: 31 December 2026

EHEDG hereby declares that the product

continuous level sensor type NSL, level detector with integrated temperature sensor type NFP-41 and point level sensor type NVS, all with G 1/2" thread (CLEANadapt "G1/2" hygienic Code "S00") made of stainless steel with black PEEK gasket

from

Negele Messtechnik GmbH, Raiffeisenweg 7, 87743 Egg an der Günz, Germany

has/have been evaluated for compliance and meets/meet the current criteria for Hygienic Equipment Design of closed process applications of the EHEDG

Certificate No. EHEDG-C2500012

Signed _____ *Hein Timmerman* _____ *President EHEDG*

Signed _____ *Karlijn Faber* _____ *EHEDG Certification Officer*

EHEDG
Karspeldreef 8
1101 CJ Amsterdam
Netherlands

©EHEDG



APPENDIX 3 EHEDG Certification – Equipment Evaluation Form

Design Evaluation Date: 20.11.2024

EHEDG File Number: EHEDG-R2400080

Certification Type: EL CLASS I

Applicant: Negele Messtechnik GmbH

Equipment: continuous level sensor type NSL, level detector with integrated temperature sensor type NFP-41 and point level sensor type NVS, all with G 1/2" thread (CLEANadapt "G1/2" hygienic Code "S00") made of stainless steel with black PEEK gasket

Other essential identification: type NSL-M, NSL-F, NSL-FR / NFP-41 / NVS-141, NVS-143, NVS-161, NVS-163, NVS-181, NVS-183

Evaluated by:

Name: Dr. Jürgen Hofmann

Date, Signature: 26.03.2025

Approved by:

Name: Yu Ming Chen

Date, Signature:

The use of the EHEDG Certification logo is justified based on the results of the design evaluation, inspection, and testing (as applicable) of the equipment for compliance with the current EHEDG Hygienic Design Criteria (HDC):

Criteria	Certification for use in Closed Processes
<input type="checkbox"/>	The equipment complies with all applicable HDC in the Guidelines.
<input checked="" type="checkbox"/>	Evidence for compliance required and provided by in-place cleanability test method according to EHEDG Doc. 2.
<input type="checkbox"/>	Evidence for compliance required and provided by in-place cleanability test method according to EHEDG Doc. 2, in-place sterilisability test method according to EHEDG Doc. 5, and bacteria tightness test according to EHEDG Doc. 7 for EL ASEPTIC Certification.
Criteria	Certification for use in Open Processes
<input type="checkbox"/>	The equipment complies with all applicable HDC in the Guidelines.
<input type="checkbox"/>	Evidence for compliance required and provided by OPC cleanability test method according to EHEDG Doc. 57.

APPENDIX 3

No.	Description
1.	EHEDG Certificate of Compliance
2.	Contract to use the EHEDG Certification Logo for equipment
3.	Appendix 1: Equipment intended for cleaning-in-place with liquids without dismantling
4.	Appendix 2: conditions for use of the EHEDG Certification Logo
5.	Appendix 3: Equipment evaluation form
6.	Evaluation report of the design of the continuous level sensor type NSL, level detector with integrated temperature sensor type NFP-41 and point level sensor type NVS, all with G 1/2" thread (CLEANadapt "G1/2" hygienic Code "S00") made of stainless steel with black PEEK gasket, no. 611TUM2025
7.	Drawings of the continuous level sensor type NSL, level detector with integrated temperature sensor type NFP-41 and point level sensor type NVS, all with G 1/2" thread (CLEANadapt "G1/2" hygienic Code "S00") made of stainless steel with black PEEK gasket, drawing nos.72xx NSL-x 6 mm EHEDG, 7114.1000-EHEDG, 7114.3200-EHEDG, 7314.1010-EHEDG, B53.720021, B53.71131x-D; original stamped
8.	Test report of the in-place cleanability test method according to Doc. 2, Test no. 611/07.11.2017.
9.	Cleaning and Installation manual nos.50144/1.2.3/2024-08-20/AS and 50000/5.2/2025-01-09/MH/EU supplied by the manufacturer
10.	Example of EHEDG Certified Logo Type EL CLASS I