

# CERTIFICATE OF COMPLIANCE



**EL Class I**

*Date of issue: 22 April 2025*

*Valid until: 31 December 2026*

*EHEDG hereby declares that the product*

***pH-sensor Memosens, type CPS61E and OCPS61E with order Code 'LC'***

*from*

*Endress+Hauser Conducta GmbH+Co. KG, Dieselstrasse 24 , 70839 Gerlingen, Germany*

*has/have been evaluated for compliance and meets/meet the current criteria for  
Hygienic Equipment Design of closed process applications of the EHEDG*

***Certificate No. EHEDG-C2500013***

Signed \_\_\_\_\_ *Hein Timmerman* \_\_\_\_\_ *President EHEDG*

Signed \_\_\_\_\_ *Karlijn Faber* \_\_\_\_\_ *EHEDG Certification Officer*

*EHEDG  
Karspeldreef 8  
1101 CJ Amsterdam  
Netherlands*

©EHEDG



## APPENDIX 3 EHEDG Certification – Equipment Evaluation Form

Design Evaluation Date: 25.02.2025

EHEDG File Number: EHEDG-R2500015

Certification Type: EL CLASS I

Applicant: Endress + Hauser Conducta GmbH+Co. KG

Equipment: pH-sensor Memosens, type CPS61E and OCPS61E with order Code 'LC'

Other essential identification:

**Evaluated by:**

Name: Dr. Jürgen Hofmann

Date, Signature: 07.04.2025

**Approved by:**

Name: Mark Morgan

Date, Signature:

4/17/25 Mark Morgan

The use of the EHEDG Certification logo is justified based on the results of the design evaluation, inspection, and testing (as applicable) of the equipment for compliance with the current EHEDG Hygienic Design Criteria (HDC):

Criteria	Certification for <b>use in Closed Processes</b>
<input type="checkbox"/>	The equipment complies with all applicable HDC in the Guidelines.
<input checked="" type="checkbox"/>	Evidence for compliance required and provided by in-place cleanability test method according to EHEDG Doc. 2.
<input type="checkbox"/>	Evidence for compliance required and provided by in-place cleanability test method according to EHEDG Doc. 2, in-place sterilisability test method according to EHEDG Doc. 5, and bacteria tightness test according to EHEDG Doc. 7 for <b>EL ASEPTIC</b> Certification.
Criteria	Certification for <b>use in Open Processes</b>
<input type="checkbox"/>	The equipment complies with all applicable HDC in the Guidelines.
<input type="checkbox"/>	Evidence for compliance required and provided by OPC cleanability test method according to EHEDG Doc. 57.

## APPENDIX 3

No.	Description
1.	EHEDG Certificate of Compliance
2.	Contract to use the EHEDG Certification Logo for equipment
3.	Appendix 1: Equipment intended for cleaning-in-place with liquids without dismantling
4.	Appendix 2: conditions for use of the EHEDG Certification Logo
5.	Appendix 3: Equipment evaluation form
6.	Evaluation report of the design of the pH-sensor Memosens, type CPS61E and OCPS61E with order Code 'LC', no. 595TUM2025
7.	Drawings of the pH-sensor Memosens, type CPS61E and OCPS61E with order Code 'LC', drawing nos. 961007985 (page 1 + 2); original stamped
8.	Test report of the in-place cleanability test method according to Doc. 2, Test no. 594+595/21.08.2017
9.	Cleaning and Installation manual nos. TI01566C/07/EN/02.21 and SD02751C/07/EN/02.23-00 supplied by the manufacturer
10.	Example of EHEDG Certified Logo Type EL CLASS I